



MISSOURI DEPARTMENT OF HEALTH
STATE PUBLIC HEALTH LABORATORY
DATAMASTER MAINTENANCE REPORT

RECEIVED
By Carol Day at 12:20 pm, Mar 08, 2010

Complete this report in duplicate at the time of the regular monthly preventive maintenance check, and whenever instrument is repaired. Send copy to Department of Health; retain original in department file.

DATAMASTER SN 204188	INV# 127308	DATE OF INSPECTION 03-08-2010
LOCATION OF INSTRUMENT (STREET AND CITY) ST. CLAIR COUNTY JAIL, OSCEOLA, MO.		TIME OF INSPECTION 1339 HRS

CHECKLIST: Place a check (✓) to the left of each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unchecked items must be corrected before using instrument.

☒ DIAGNOSTIC CHECK (PRINTOUT ATTACHED)

☒ COMPUTER

☒ DETECTOR

☒ PROGRAM

☒ FILTERS

☒ HEATERS SAMPLE CHAMBER **50 °C**

☒ QUARTZ STANDARD

☒ FLOW DETECTOR

☒ CALIBRATION

☒ PUMP HIGH SPEED

☒ PRINTER

☒ INDICATOR LIGHTS

☒ TIME AND DATE

☒ SIMULATOR TEMPERATURE (34 °C ± 0.2°C) **(34.1°C)**

☒ CALIBRATION CHECK -

Run three tests using a standard solution. All three tests must be within ± 5% of the standard value and must have a spread of .005 or less. Check the box corresponding to the standard solution being used. (PRINTOUT ATTACHED) (USE RECIRCULATION PUMP)

☒ 0.100% STANDARD - MUST READ BETWEEN 0.095% and 0.105% INCLUSIVE

☐ 0.040% STANDARD - MUST READ BETWEEN 0.038% and 0.042% INCLUSIVE

(ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

TEST 1 ✓ .098	TEST 2 ✓ .099	TEST 3 ✓ .099
-----------------------------	-----------------------------	-----------------------------

☒ PERFORM R.F.I. TEST (PRINTOUT ATTACHED)

☒ NUMBER OF REFUSALS, SINCE LAST MAINTENANCE REPORT, AND NUMBER OF BREATH TESTS IN EACH RANGE AS FOLLOWS (DO NOT INCLUDE SIMULATOR TESTS)

REFUSALS 0	(0-.04) 5	(.05-.09) 2	(.10-.14) 0	(.15-.19) 0	(Over .19) 1
-------------------	------------------	--------------------	--------------------	--------------------	---------------------

List any new parts and describe any alteration or modification that was made to restore the instrument to operate satisfactorily and within established limits (use other side if necessary)

THIS INSTRUMENT IS OPERATING WITHIN DHS SPECIFICATIONS

SOLUTION: REPCO MARKETING INC. LOT# 09001 .10% SOLUTION
EXPIRES 4-22-11

INSPECTING OFFICER

SIGNATURE W.M. Bush	PRINT NAME W.M. BUSH
TYPE II PERMIT NUMBER/EXPIRATION DATE 920004 1-15-11	TELEPHONE NUMBER 417-895-6868

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 09001

EXPIRATION DATE: April 22, 2011 at 11:59 p.m.

RepCo Marketing, Inc. certifies the following:


RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 09001 of Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .1214 gms/dl \pm .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 \pm 3% gms/210L Breath when heated to 34 Degrees Celsius \pm 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is April 23, 2009.
The expiration date for this lot number is April 22, 2011
at 11:59 p.m.

This document is a true representation of the original Certificate of Analysis.



Cecil B. Garner, President
RepCo Marketing, Inc.

State of Missouri
DEPARTMENT OF HEALTH



P E R M I T
TYPE II



W. MITCHELL BUSH

is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air. Issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 01/15/09
Number 920004
Expires 01/15/2011

MO 590-0771 (7-88)

John J. Mathewson

Director of State Public Health Laboratory

Margaret T. Donnelly

Director, Department of Health

Lab. 4 (R7-88)

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
DRC DIAGNOSTIC SERIAL NUMBER 204186
03/08/00

ARGENT LINE: 10100

SUBJECT NAME:

PRISON

DOB: 09/09/67 SEX: M

STREET: 10100

ASSISTING OFFICER:

BUSH/006

OFFICER: 10100

TELEPHONE OFFICER:

BUSH/006

OFFICER: 10100

OFFICER: 10100

EXPIRATION DATE: 01/01/01

RECEIVED: 01/01/01

--- BREATH ANALYSIS ---

RECEIVED: 01/01/01

Operator Signature

WM B

2

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
DRC DIAGNOSTIC SERIAL NUMBER 204186
03/08/00
13:59

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM (04-07-0009): OKAY

HEATERS: OKAY

SAMPLE CYCLING: OKAY

FLOW DETECTOR: OKAY

PUMP: OKAY

HIGH SPEED: OKAY

DETECTORS: OKAY

FILTERS: OKAY

QUALITY STANDARD: OKAY

CALIBRATION: OKAY

PRINTED: 03/08/00

10100 (01/01/01) - 01/01/01 (01/01/01) - 01/01/01 (01/01/01)
RECEIVED: 01/01/01 (01/01/01) - 01/01/01 (01/01/01)
COMPLIANCE: 01/01/01

Operator Signature

WM B

2208-02

Face This Side Down - This Edge In First

BAC DataMaster Evidence Ticket

MISSOURI STATE HIGHWAY PATROL
BAC BREATHALYZER SERIAL NUMBER 294186
03/05/10

TESTING OFFICER:
DUSH-008
OFFICER ID: 1111 250
PERMIT NUMBER: 920004
EXPIRATION DATE: 01/15/11
MODEL (BAC) 0070
MODEL (MANTENANCE)
ST. LOUIS COUNTY

--- SUPERVISOR NAME ---

BLIND TEST	.000	13:47
INTERNAL STANDARD	VERIFIED	13:47
EXTERNAL STANDARD	.000	13:47
BLIND TEST	.000	13:48
EXTERNAL STANDARD	.000	13:49
BLIND TEST	.000	13:49
EXTERNAL STANDARD	.000	13:50
BLIND TEST	.000	13:51

P - 2
FOL - 1
FAL - 10000

Operator Signature

